


What is the hadron-therapy?

Use of ions for the radiotherapeutic treatment of tumours



So we can answer to the question:
Why clinical proton beams?

- penetration depth is ***well-defined*** and ***adjustable***
- most energy at ***end-of -range***
- protons travel in ***straight lines***
- dose to ***normal tissue*** minimised
- ***no dose*** beyond target

PROTONS PERMIT TO DELIVER A HIGH DOSE TO
THE TUMOUR SPARING THE SURROUNDING TISSUES

The Laboratori Nazionali del Sud



In Catania we developed a facility **CATANA** for the treatment of ocular tumours with proton beams of 62 AMeV






**Treatment of the
choroidal and iris
melanoma**


**In Italy about 300
new cases for year**

LNS

**Superconducting
Cyclotron is the
unique machine in
Italy and South
Europe used for
protontherapy**





Laboratori Nazionali del Sud –INFN Catania, Italy




PRESENT TREATMENT ROOM

- 0 ° respect the switching magnet
- 80 meter after extraction
- 3 m proton beam line






Surgical Phase (Tantalum clips insertions)



CLIPS: characterize position and size of tumor volume







EYEPLAN

In origin developed by Michael Goiten e Tom Miller (Massachusetts General Hospital) e ora maintained by Martin Sheen (Clatterbridge Center for Oncology) e Charlie Perrett (PSI)

Treatment Planning System Output

Isodoses curves for different planes



Patiens look at the fixation light
during the treatment




ROTON BEAM





15 03 2002 10:58

Patient Distribution by Origin Region




Patient Distribution

Total number of patients : 66



Mean age: 57.6 yrs



Why to start a Simulation Work ?

Therapy with hadrons still represents a pioneering technique

Today the development of a hadron-therapy facility requires a long experimental work due to the lack of
SIMULATION TOOLS

Our work is inserted in the more general
medical-physics GEANT4 activity
and represents just a different application of a more general approach in the medical-physics field

Why to start a Simulation Work ?

This work concerns mainly:

Design and optimization of the
transport beam line elements:

Test of the elements

Test of the detectors

Reconstruction of the dose
distributions:

To measure dose distribution
also in difficult experimental
region

To verify the radiotherapy
treatment planning systems

Why to start a Simulation Work ?

So we start our simulation work using GEANT4:

- To simulate our complete beam line with all its elements and
- To reproduce all the dose distributions

It's impossible to conceive a modern detector w/o simulation

Rossi and Greisen 1941, Rev. Mod. Phys. 13:240

Our GEANT4 Application:

hadronTherapy.cc

Complete simulation of CATANA
hadron-therapy beam line with two dosimeters


- Depth Dose Distribution in Water (Bragg curve):
Markus type ionization chamber;
- Lateral Dose Distribution:
Radiochromic film;

Material and Methods

DETECTORS USED FOR DOSE DISTRIBUTION MEASUREMENTS

DEPTH DOSE DISTRIBUTION

- Markus Ionization chamber



LATERAL DOSE DISTRIBUTION

- GAF Chromic Film




Markus Chamber layout

Irradiated GAF Chromic

Resolution 100 µm for DDP and 200 µm for LDP


IPICI RIVELATORI PER LA RICOSTRUZIONE DELLE CURVE DI DOSE



Camera a ionizzazione



TLD




Gaf




Diodi



Diamanti




MOSFET




Scintillatore + CCD

Experimental 'PURE' Bragg curve




DETECTOR	PEAK DEAPTH	PEAK – PLATEAU RATIO	F.W.H.M.	Distal - dose falloff $d_{80\%-20\%}$	Practical Range ($d_{10\%}$, ICRU 59)
MARKUS	30.14	4.68	3.19	0.50	31.15

Experimental 'modulated' Bragg curve




Experimental Lateral Dose Distribution



	$R_{90\%-50\%}$	P_s [mm]	P_d [mm]	Symmetry [%]	Homogeneity at 95% level [mm]
LNS	0.92	0.8	1	2.4	21
CCO	0.93	0.75	0.75	2.6	23

Bragg Curve Reconstruction

Detector is simulated with
20 K air cylindrical slices, 200 μm thick
to reproduce experimental Markus chamber response





Energy deposited in each slice is collected

We calculated range values for the detector simulation validation from Bragg curve

Bragg Curve Reconstruction



Water box with ionisation chamber





Water box + detector
for Bragg curve as
simulated

Validation of Simulated Detector (From a Software point view)

Comparison with NIST ranges data





NIST DATABASE COMPARISON



Beam Line Simulation: Scattering system

DOUBLE SCATTERER FOIL WITH CENTRAL STOPPER


15 μm + 25 μm + 7 mm thick copper beam stopper



Permits to obtain an homogeneous lateral dose distribution at isocenter



Real hadron-therapy
beam line




GEANT4 simulation


Each element of the line can be modified (in shape, material and position) and other kinds of dosimeters can be easily inserted

Physics models: comparison with experimental data

Standard Processes



Standard + hadronic



Kolmogorov test

process	P-value	Test
---------	---------	------

Standard.

0.069

OK

Standard + Had.

0.40

OK

Low Energy

0.51



OK

Low En. + Had

0.699

OK

Low E




The best hadronic model we find in our investigation is the precompound (*Theory Driven Model* of GEANT4)

Validity range up to 6.8 MeV for protons and for any target material (Z>5)

Hadronic processes


process	class
Inel. Nucl. Scatt..	G4HEProtonInelastic
Elastic Nucl. Scatt.	G4HadronElasticProcess
Fission	G4LFission

Experimental data comparison




Low energy libraries and hadronic physics

Bragg peaks at different energies




Experimental Data Comparison




Simulated and experimental range in water and copper

Kolmogorov test

materiale	P-value	test
Acqua	0.9876	Accettato
Rame	0.999	Accettato



Lateral Distribution: comparison with experimental data




Monte Carlo examples
to study and
improve the
transport beam line


H.Paganetti, Monte Carlo method to study the proton fluence
for treatment planning
Med. Phys. 25, 1998

Study of the influence
in the dose distribution

PHASE SPACE DISTRIBUTION
reported versus the dose
plan transversal to the beam direction.






Phase space distribution along the transport beam line in Catania





Lateral distribution in the phase space distribution

Initial beam





Isocenter
(patient location)




Lateral distribution in the phase space distribution

A variation of a beam element along beam line can be studied with Monte Carlo studying phase space distributions: conventional dosimeter cannot be so sensible



Dose distributions inside patients




For our proton beam line penumbra increases up to 34%


Energy sigma increases up to 83%



Modulator s
reconstructi



Geant4 Modulator we
simulated next week



Monte Carlo in the clinical practice can be limited by
the long calculation times

On the other hand *velocity* is mandatory for a medical
physicis and a medical doctor when they are to plan a
treatment

*GRID technologies could represent a solution
see the talk of Giorgio Russo*

eature Developments



DICOM images insertion




More realistic dose distributions



Statistical comparison between
analytical based treatment
planning systems

We decided to start this work with the Clatterbridge center for oncology where the EYEPLAN TPS system was developed and now maintained





Low Energy range measurements (5 MeV – 100 KeV)




We are studying different Hadronic models and the nuclear components to the released dose for proton and carbon




In the next week we'll compare a bragg peak for a 62 MeV Carbon beam with GEANT4 simulation


The Sicilian Center for Hadron Therapy



*250 MeV for protons, 1500 MeV for Lithium
and 3000 MeV for Carbon Ions*



Design di hadronTherapy





hadronTherapy design

March 8, 2003 G.A.P. Cirrone

Caratteristiche del fascio primario simulato

Sulla base dei dati sperimentali e delle caratteristiche dell'acceleratore superconduttore:




Energia Iniziale	63.4 MeV
δE	300 keV
σ Posizione	1 mm
σ Momento	0.0001

Caratteristiche del fascio primario simulato

Sulla base delle curve di Bragg sperimentali abbiamo definito le caratteristiche iniziali del fascio di protoni

- Energia
- *Spread* in energia
- Distribuzione spaziale
- Distribuzione dei momenti

“Fotografia” del fascio all’isocentro




La terapia con adroni è considerata ancora oggi una tecnica pionieristica

Numerosi centri in Europa si svilupperanno nei prossimi anni:


- RPTC, Munich Germania
- PSI, Villigen Svizzera
- Trento, Italia
- TERA, Italia
- Heidelberg/GSI, Darmstadt Germania
- Centro di Adroterapia, Catania Italia

Perche GEANT4 in applicazioni mediche

- Codice Trasparente (open source)
- Librerie specifiche per i processi fisici alle basse energie e per i processi adronici
- Tecnologia Object Oriented
- Numerose validazioni del software
- Supporto On-Line
- Interfaccie User Friendly
- Interattività



Validazione sui dati sperimentali: matrice degli indici γ





Distribuzione delle iso-gamma nelle regioni per le quali il test fallisce

Validazione del programma sviluppato





Confronto dei parametri di interesse medico con i dati sperimentali.




Curva di Bragg

- Range Pratico
- Rapporto picco-plateau
- Distribuzione spaziale del fascio di protoni
- Penombra Laterale
- FWHM
- $R_{90/50}$

Dati sperimentali: distribuzioni di dose in profondità



Elementi tipici di una linea di protonterapia




1. Primo diffusore
2. Secondo diffusore, stopper di ottone
3. Camera con il modulatore
4. Collimatore
5. Camere Monitor
6. Collimatore finale e del paziente
7. Cubo contenente il rivelatore sensibile

Sistema di diffusione

Maggiore trasparenza rispetto agli altri codici
nei processi fisici implementati

Processi fisici implementati In hadronTherapy



1 PeV – 1 keV per fotoni ed elettroni
1 PeV – 10 keV per protoni e ioni

100 GeV – 250 eV per fotoni ed elettroni
100 GeV – 1 keV per protoni e ioni